



Patron  
Lt. Gen. J. R. Bhardwaj  
Past President, IAPM

Secretary General APP,  
Chief Coordinator &  
Organizer VILAC  
Dr. Rakesh Sharma  
North Delhi Pathology Clinic  
Cell. : 9873247824

Chairman,  
Advisory Body, VILAC  
Prof. S K Sood

President  
Dr. Sunil Dargar  
Cell : 9811078381

Secretary  
Dr. Subodh Gupta  
Ph. : 22751743, 65416091  
Cell : 9810128454

Treasurer  
Dr. Neeraj Jain  
Cell : 9313679026

Advisors  
Prof. A K Agarwal  
Prof. K B Logani  
Prof. P K Shukla

Exec Vice President  
Dr. Alok Jain  
Cell : 9811364458

Vice Presidents  
Dr. Navin Dang  
Dr. Juhi Chandra  
Dr. Vibha Malik  
Dr. Savita Nagpal  
J K Arora

Executive Members  
Dr. J S Suri  
Dr. Rajan Verma  
Dr. G S Agarwal  
Dr. B C Gupta  
Dr. A. P. Aggarwal  
Dr. Vagish Dave  
Dr. Gajender Yadav  
Dr. Mohit Sharma  
Dr. Prashant Gupta  
Dr. Neha Trivedi Kaushik  
Dr. Anil Kumar Gupta  
Dr. Sanjay Gupta  
Dr. Rajesh Makashir  
Dr. Mudit Agarwal  
Dr. M K Wadhwa



ISO 9001:2008 CERTIFIED ORGANISATION

Continuous Quality Improvement



## VILAC 2011 - November 24<sup>th</sup>, 2011

### Statistical Analysis Cycle-XI

Nov 16<sup>th</sup> - Nov 23<sup>rd</sup>, 2011

Participant Dr. Angeli Misra of Lifeline Laboratory Code: V-016

Parameter	Units	Group Mean	S.D.	CV %	Your Value	SDI
<b>Biochemistry</b>						
Urea	mg/dL	275	30.7	11.2	294	0.6
Creatinine	mg/dL	4.0	0.4	10.0	4.2	0.4
Uric Acid	mg/dL	8.7	0.7	8.0	8.4	0.4
Cholesterol	mg/dL	200	16.2	8.1	205	0.3
HDL Chol	mg/dL	58	9.4	16.2	52	0.6
Triglycerides	mg/dL	189	16.8	8.9	192	0.2
Bilirubin	mg/dL	2.8	0.4	14.3	2.8	0.0
SGOT	U/L	106	10.5	9.9	114	0.8
Calcium	mg/dL	9.9	0.7	7.1	10.5	0.9
Inor Phos	mg/dL	6.3	0.9	14.3	7.0	0.8
Protein	g/dL	6.1	0.6	9.8	5.6	0.8
Glucose	mg/dL	181	14.9	8.2	192	0.7
<b>Hematology</b>						
HB	g/dL	13.0	0.8	6.2	13.1	0.1
TLC	per cumm	7421	470	6.3	8020	1.3
TRBC	mill/cumm	4.2	0.3	7.1	4.3	0.3
Platelet	thous/cumm	317	28.6	9.0	302	0.5
Blood Group	ABO	B	--	--	B	--
Rh	Pos/Neg	Positive	--	--	Positive	--
<b>Immunoassays / Special Tests - Participation in at least 3 parameters Mandatory</b>						
HbA1C	%	5.2	0.7	13.5	4.7	0.7
HbA2	%	3.1	0.2	6.5	3.2	0.5
TSH	µIU/mL	1.4	0.2	14.3	1.2	1.0
Prolactin	ng/mL	6.9	1.0	14.5	6.9	0.0
Testosterone	ng/mL	2.5	0.4	16.0	2.1	1.0
PSA	ng/mL	2.6	0.4	15.4	3.0	0.9

• Please inform about any transcription error with in 48 hours.

• NR= Not Reported

Interpretation of SDI:

SDI Value	0.0 - 0.5	0.5 - 1.0	1.0 - 2.0	2.0 - 3.0	3.0 and above
Interpretation	Excellent	Good	Acceptable	Needs Improvement	Urgent action required

*Rakesh Sharma*  
Dr. Rakesh Sharma  
Chief Coordinator & Organizer VILAC

To,  
Dr. Angeli Misra  
Lifeline Laboratory  
H-11, Green Park Extn.  
New Delhi 110016  
M: 9811392286

Prepared by *[Signature]* Checked by *[Signature]*

*[Signature]*  
26/11/11